

Administration of Daman & Diu (UT).
Directorate of Medical & Health Services,
Primary Health Centre,
Moti Daman, Daman.

No. DMHS/DD/EQUIPMENT /2009-10/3292

Date : 09/02/2010.

LIMITED TENDER NOTICE

The Director. Medical & Health Services, Daman and Diu, Daman invited sealed limited tender from registered manufacturers and various authorised dealers for below mentioned items for Cardiac cum Trauma Care Ambulance under "M.P. LAD Scheme".

Sr. No.	Particulars
1.	Roll in stretcher cum Trolley model
2.	Patient restraint system
3.	Porta Vac Aspirator – Electronic Suction unit
4.	Scoop Stretcher
5.	Folding stretcher
6.	Portable Ventilator
7.	Defibrillator / Monitor
8.	Multipara monitor with 3 lead ECG, SPO2, Pulse, NIBP.
9.	Oxygen therapy mask, Ambu bag with face mask
10.	Laryngoscope set along with cells
11.	Neck immobilizer (Hard Cervical Collar, Small & Medium)
12.	Splints inflatable for immobilizer of limbs.
13.	Oxygen cylinder
14.	Emergency kit
15.	Spine board
16.	Fracture splint kit
17.	Syringe infusion pump (Infutek 400)

The tender form along with terms and conditions can be had on payment of Rs.250/- (non-refundable) from the Directorate of Medical & Health Services, Primary Health Centre, Fort Area, Moti Daman – 396 220 during the working hours up to 22.02.2010 at 11.00 am. The Earnest Money Deposit of Rs.25,000/- in form of demand draft or FDR of any scheduled bank in favour of the Director, Medical & Health Services, payable at Daman is to be submitted along with Technical Bid. The Technical Bid which is not accompanied by EMD shall be summarily rejected.

The tender is to be submitted in two bid system, each of which is required to be submitted in separate envelope. The Technical Bid should contain all documents pertaining to technical specification including Tender Form for technical offer super scribing on the envelope as "Sealed Cover No. 1 – Technical Bid for supply of **Equipments for Cardiac cum Trauma Ambulance**", and Sealed Cover No. 2 - Financial Bid for supply of **Equipments for Cardiac cum Trauma Ambulance**".

The tender should reach to this Directorate on or before 22-02-2010 up to 12.30 pm. The Technical Bid will be opened on the same day, if possible at 3.00 pm. in the presences of tenderers or their authorized representative, if any. The details of the tender notice can also be downloaded from the official website www.daman.nic.in and can submit the tender along with tender fee in form of Demand Draft to be drawn in the name of Director, Medical & Health Services, Daman.

The Director, Medical & Health Services, Daman & Diu is reserved rights to accept or reject any tender in part or whole or all tenders without assigning any reason.

(Dr. K.Y. Sultan)
Director
Medical & Health Services

U.T. ADMINISTRATION OF DAMAN & DIU
DIRECTORATE OF MEDICAL AND HEALTH SERVICES,
DAMAN

CONDITIONS FOR THE SUPPLY OF EQUIPMENTS FOR CARDIAC CUM TRAUMA CARE
AMBULANCE.

Short Tender Notice No. DMHS/DD/ EQUIPMENT/2009-10/3292

Dated 09-02-2010.

1. The rate(s) quoted should be strictly for free delivery at F.O.R. Government Hospital, Daman and will be valid and operative for supply orders issued on or before 31-03-2010.
2. The rates should be quoted inclusive taxes, Central Excise and installation & commissioning charge, training etc.
3. (a) VAT / Sales Tax will be paid on the items on which it is chargeable under the VAT / Sales Tax Act or the Rules made there under. C.S.T. will be paid @ 4% against form 'D'.
(b) Request for issue of Form 'D' with supply order will not be entertained as Form 'D' will be issued along with payment of respective bill only.
4. Except C.S.T. all other Taxes/Duties/Royalties Charges payable on the sales/transport etc. within and/or outside the state of supplier shall be payable by the supplier.
5. No extra charge for packing, forwarding and insurance etc. will be paid on the rates quoted.
6. The rates should be quoted only for the items specified in the list of requirement and should be for the items of given special mark/manufacture.
7. Rates quoted for items other than the required specification/mark/manufacture may not be considered. However, indigenous manufactures may quote their own makes provided the specifications confirm to the standard(s)/requirements of the given specifications/mark/ manufacturer.
8. Where specifications/mark/manufacture are not specifying by this office, the rates should be quoted only for the 1st class and standard quality.
9. The Tenderer should specify the name of the manufacturer for the items quoted by him along with catalogue of the item.
10. The decision of the Tender Inviting Officer for acceptance/rejection of any articles supplied including the decision for equivalent specifications, standard and quality etc. of articles shall be final.
11. The Tenderer should send in advance or enclose along with technical bid an amount of Rs. 25,000-00 (Rupees Twenty Five Thousand only) as Earnest Money Deposit in form of Demand Draft / F.D.R. of any Nationalized / scheduled Bank at payable at Daman in favor of the officer inviting tenders i.e. Director, Medical & Health Services, Daman. The E.M.D. submitted other than form mentioned above will not be accepted. Tender received without Earnest Money Deposit will be summarily rejected. The Technical Bid and the Financial Bid of that firm will also considered as rejected.
12. (a) The successful Tenderer will have to pay within 10 days from the date of demand, an amount equal to 10% of the total value of articles, that may be order, as the amount of security deposit.
(b) Non-receipt of Security Deposit within stipulated time limit will result in automatic cancellation of the order for supply without any intimation.
(c) However in case if any articles are received for which the Security Deposit may not have been deposited, the full Security Deposit as may be due from the supplier will be recovered from the bill(s) for such articles.

13. The amount of Earnest Money paid by the successful Tenderer(s) will be adjusted against the amount of Security Deposit to be paid by the successful Tenderer(s) as per condition No. 12 above.
14. The tender should be neatly typed only on letterhead carry the name of supplier and the signature of the Tenderer. No overwriting, correction or erasures will be considered.
15. The amount of Earnest Money paid by the Tenderer(s) whose tenders are not accepted will be refunded to them by Cheque or Demand Draft (as may be convenient to the Tender Inviting Officer if the amount is above Rs.200/-) drawn on an branch of State of India or its subsidiary Schedule Bank. Where this mode of payment is not possible the amount will be refunded at the cost of the Tenderer.
16. Only on satisfactory completion of the supplies ordered for and on payment of all bills of the supplier, as to be admitted for payment, the amount of Security Deposit/Earnest Money will be refunded after expiry of guarantee/warranty period, if any, or any such date/period as may be mutually agreed upon.
17. The tender inviting officer will consider extension of time for remitting the Security Deposit as demanded. However, in case of denial to consider such extension the supplier is bound to abide by the limit given and liable to make good any loss to the Government on account of his failure to abide by the time limit.
18. (a) Railway Receipt or other transport document should be drawn in favor of the Officer Inviting Tenders.
(b) Railway Receipt or other transport document should not be send by V.P.P. or through any Bank as this being a Government Office it is not possible to clear cash demands of Post Office/Bank for delivery of R.R. or other transport documents unless we have agreed to it as a special arrangements.
(c) Railway Receipt or transport should be sent to this Office by Registered Post immediately on dispatch of goods from dispatching end.
19. The supplies of Medical equipments, Store equipment etc. of inferior quality standard or of different specifications, brand, manufacturer etc. other than that ordered specified and/or incomplete or broken articles will not be accepted. The supplier will have to replace the same at his own cost and risk. Intimation of non-acceptance of any stores Machinery and equipment etc. will be sent to the supplier within 10 days from the date of receipt of the stores and the same will be returned to the supplier at his own cost and risk, if he so desires and intimates accordingly within 15 days from the date of dispatch of intimation of the non-acceptance. However, if no communication is received within 15 days from the date of communication the tender Inviting Officer will not be responsible for any damage, loss etc. of such rejected articles.
20. In case of failure to replace the accepted and rejected articles from supplies made, as mentioned in the conditions, the loss undergone by the Government will be recovered from the suppliers Security Deposit/Earnest Money or payment due of any bill(s) to the extent required.
21. In case of failure to supply the store, machinery and equipment etc. ordered for, as per conditions and within the stipulated time, the same articles will be obtained from the Tenderer who offered next higher rates or from any other sources, as may be decided by the tender inviting Officer and the loss to the Government on account of such purchases(s) shall be recovered from the former suppliers Security Deposit/Earnest Money or bills payable. The suppliers shall have no any right to dispute with such procedure.
22. Extensions of time limit for supplies shall be consider by the Tender Inviting Officer. The extension so granted may be with levy of compensation for delay in execution of supply order up to 5% of the cost of supplies ordered for at the discretion of the authority competent to grant extension of time limit provided such

- request is made well in time, depending upon the circumstances and such decision in the matter will be final.
23. Demurrage charges paid by the Tender Inviting Officer on account of delayed receipt of dispatch documents intimation will be recovered from the bills payable to the supplier.
 24. (i) The supplier(s) of the machinery and equipments shall have to supply spare parts as and when required in on an agreed basis i.e. on agreed on the published catalogue/price lists for an agreed period.
(ii) The supplier (s) shall give adequate advance notice before any spare parts goes out of production to enable the tender Inviting Officer to order for spare parts required in one lot for the life time of respective machinery.
(iii) If mutually agreed to the supplier(s) shall make available the blue prints of drawing etc. of the spares if any when required in connection with main machinery/equipments.
 25. If at any time after the order for supply of Machinery/stores/equipment the Tender Inviting Officer shall for any reason whatsoever not require the whole or part of the quantity thereof as specified in the order the tender Inviting Officer shall give notice in writing of the fact to the supplier(s) who shall have no claim to any payment of compensation what so ever on account of any profit or advantage which the supplier(s) might have derived from the supply of articles in full, but which did not derive in consequence of the full quantity of articles not having been purchased, nor shall have any claim for compensation by reasons of any alterations having been made in the original instructions which shall invoice any curtailment of the supply originally contemplated.
 26. The Earnest Money(s)/Security Deposit(s) paid by the tender(s) earlier against any tender(s) or supply order(s) is/are not adjustable with Earnest Money or Security Deposit required by these conditions,
 27. All bills should be in DUPLICATE and should invariably mention the number and date of supply order.
 28. All bills for amount above Rs.5000/- should be pre-receipted on a Revenue Stamp of proper value. Bills for amount exceeding Rs.5000/- not pre-receipted on Revenue Stamp of proper value will not be accepted for payment.
 29. Each bill in which Sales Tax is charged must contain the following certificate on the body of the bill.
"CERTIFIED that the goods on which Sales Tax has been charged have not been exempted under the Central Sale Tax Act or the Rules made there under and the amount charged on account of Sales Tax on these goods is not more than what is payable under the provisions of relevant Act or Rules made there under".
 30. The Purchase Committee will open the Tenders in presence of tenderers or their representatives, if any present in the Office of the Tender Inviting Officer. The Purchase Committee will first open the technical bid and considering the technical specification of Hospital equipment and furniture. The Committee will open the financial bid only of those firms who are qualified for technical bid as per minimum specifications given by the department.
 31. The right to accept or reject without assigning any reasons any or all tenders in part or whole is reserved with the Tender Inviting Officer and his decision(s) on all matters relating to acceptance or rejection of the tenders as a whole or in part will be final and binding to all.
 32. The tenders/offers received do not confirm with the terms and conditions of this office will be summarily rejected. If any firm desires to consider exemption from payment of Earnest Money Deposit, certified copies of its Registration with D.G.S. & D. should be attached to their tenders.

33. If the tenderers whose tender is accepted, fails to execute the supply order within stipulated time the Earnest Money Deposit of such tenders will stand forfeited to the Government.
34. In case, the supplier does not execute the supply order placed with him, the EMD of the supplier will be forfeited to the Government and the contract for the supply shall terminated with no further liabilities on either party to the contract.
35. No separate agreement will be required to be signed by the successful tender(s) for the purpose of this contract for supply. Rates tendered/offered in response to the concerned Tender Notice shall be considered as acceptance of all above terms and conditions for supply for all legal purpose.
36. The items as mentioned in the list are the approximate estimates invited and actual purchase may more. Accordingly the successful Tenderer has no right for any loss/damages with reference to approximate requirement shown in tender and actual requirement.
37. The tender will be accepted during working hours up to 22-02-2010 at 12.30 p.m. and opened on same day at 3.00 pm, if possible in the office of the Director, Medical & Health Services, Primary Health Center Moti Daman in the presence of the Purchase Committee and Tenderer(s) or their representative(s) if present.
38. Supplier may ensure the goods at his own cost to safeguard the delivery of such goods dispatched by him to the consignee, the department will not be responsible for the damage or pilferage of goods during transit.
39. The tenderers should attached details of civil work to be carried out by this Department for installation/commissioning of Incinerators.
40. The tenderers should give the guaranty/warranty for the period of not less than one year from the date of installation against any manufacturing defect.
41. The successful tenderers should install and commissioning the equipment and furniture at the site suggested by the office.
42. The successful tenderers shall be bound to provide training if any required without any extra charges during commissioning.
43. The tenderers shall be bound to give assurance for undertaking the annual maintenance contract after the expiry of guarantee/warranty period.
44. The tenderers must have to submit the tender in two bid cover system, 1st cover of Technical bid and 2nd cover of financial bid and both separate cover should be in one sealed cover mentioning "TENDER FOR RQUIPMENTS FOR CARDIAC CUM TRAUMA CARE AMBULANCE" and due date on cover.

Signature & Designation of
Tender Inviting Officer

Sd/-
(Dr.K. Y. Sultan)
Director
Medical & Health Services
Daman

The above terms and conditions are accepted and are binding to me/us.

Place: Signature of tenderers
Dated: Name of tenderers with seal
of the firm

NOTE: Please return one copy of these terms & conditions dully sign with seal of firm along with the tender.

Administration of Daman and Diu (UT)
Directorate of Medical & Health Services,
Primary Health Centre, Moti Daman, Daman.

TENDER FORM (FINANCIAL BID)

Ref: TENDER NOTICE NO.DMHS/DD/ EQUIPMENTS /2009-10/3292
dated 09/02/2010 for supply of equipments for Cardiac cum Trauma
Care Ambulance.

Sr. No.	Particulars	Unit	Rates per unit
EQUIPMENTS			
1.	Roll in stretcher cum Trolley model	One	
2.	Patient restraint system	One	
3.	Porta Vac Aspirator – Electronic Suction unit	One	
4.	Scoop Stretcher	One	
5.	Folding stretcher	One	
6.	Portable Ventilator	One	
7.	Defibrillator / Monitor	One	
8.	Multipara monitor with 3 lead ECG, SPO2, Pulse, NICP.	One	
9.	Oxygen therapy mask, Ambo bag with face mask	One	
10.	Laryngoscope set along with cells	One	
11.	Neck immobilizer (Hard Cervical Collar, Small & Medium)	One	
12.	Splints inflatable for immobilizer of limbs.	One	
13.	Oxygen cylinder	One	
14.	Emergency kit	One	
15.	Spine board	One	
16.	Fracture splint kit	One	
17.	Syringe infusion pump (Infutek 400)	One	

Seal

Signature and Seal of Tenderer

Date

Administration of Daman and Diu (UT)
 Directorate of Medical & Health Services,
 Primary Health Centre, Moti Daman, Daman.

TENDER FORM (TECHNICAL BID)

Ref: TENDER NOTICE NO.DMHS/DD/ EQUIPMENTS /2009-10/3292
dated 09/02/2010 for supply of equipments for Cardiac cum Trauma Care Ambulance.

Sr. No.	Particulars	REMARKS OF TENDERER TO PROVIDE EQUIPMENTS AS PER SPECIFICATION "YES" or "NO"
EQUIPMENTS & ITS SPECIFICATION		
1.	<p>Roll in stretcher cum Trolley model:-</p> <p>Self loading stretcher trolley, with easy loading mechanism. It should operate by applying the two handles fitted left and right at the foot end of the under construction this trolley's legs will fold away neatly when it is pushed into the ambulance. They automatically unfold again when pulled out of the ambulance and lock into position. The adjustable backrest can be locked in different positions to give the patient extra comfort. Fitted with 4 wheels, 2 of which are fixed, 2 are swivel wheels to achieve better maneuverability. It is also equipped with 2 fold down side rails to allow for easy patient transfer. The frame should be of anodized aluminum, with polyurethane and polypropylene based patient bed, which is joint less, resistant to all body fluids. It should be 10G Crash Test approved, CEN 1789 compliant. The wheels should be of dia 200 mm or more for better stability.</p>	
2.	<p>Patient restraint system :-</p> <p>It is designed to provide full body immobilization. It is compatible with all the full body length back boards with eight or more hand holds or strap slots. It should be CE approved.</p>	
3.	<p>Porta Vac Aspirator – Electronic Suction unit :-</p> <p>Emergency Aspirator Independent of any power source, the suction pump should be simple to operate by foot, hand or knee. It may be used to clear patients airway safely and efficiently, anytime, anywhere. All components are easily accessible, may be removed, cleaned and replaced. Completely autoclavable. The aspirator is complete with aspiration pump complete, suction vessel and over spill protection. Sizes - 223X165X113 mm., Weight - 1.4 kg., Suction tube diameter - 10 mm., Vacuum - 0.7 bar/10 psi., Suction Capacity - 320 mm., Suction Rate - 25 l/min.</p>	
4.	<p>Scoop Stretcher :-</p> <p>The Spencer scoop stretcher telescopes to accommodate the tallest patient and folds for compact storage. Designed so that attendants can uncouple either end of the stretcher and gently scoop up the patient using scissors. Type closing motion. Tubular aluminum construction with extruded aluminum blades. Includes n. 3 nylon restraint straps.</p> <ul style="list-style-type: none"> - Minimum Length 167 cm, - Maximum Length 201 cm - Folded 120 cm - Width 42.5 cm - Depth Folded 9 cm - Weight 10 kg - Load capacity 170 kg - Wall Bracket - Carrying Bag 	

5.	<p>Folding stretcher :-</p> <p>Emergency Pole Stretcher (Four Folding) Folds lengthwise and crossbar for compact storage with non-slip handgrips. Four feet rough terrain. Cross bars add comfort and support for patients.</p>	
6.	<p>Portable Ventilator :-</p> <p>Application : adult and child, Ventilation modes : controlled mechanical ventilation or assisted controlled mechanical ventilation with built in positive and expiratory pressure, with either pure oxygen or air/oxygen mixture, spontaneous ventilation with pressure support and positive end expiratory pressure applicable to non-invasive ventilation at the mask with pressure controlled safety mode minimum respiratory rate adjustment to guarantee sufficient ventilation even in the event of a half in spontaneous breathing by the patient. Performance characteristics : Tidal volume :100 ml to 1500 ml. Respiratory rates : 6 to 40bpm PEEP : 0 to 15 cm H₂o Pressure support : 0 to 44 cm H₂o Inspiratory trigger : 0.5 to – 4 cm H₂o I/E ratio : 1/3 to 1/1 Fio₂: 100% or air / o₂-mixture. CMV-ACMV position 100% CMV-ACMV position air/o₂. PS-PEEP position air/o₂. Alarms : Alarms audible and visual with LCD alarm message display. P_{maxi} : 10 to 80 cm H₂o P_{mini} : 5 to 10 cm H₂o above peep F_{maxi} : 20 to 60bpm. Patient disconnection power supply failures. Monitoring : Bar graph pressure display : 4 to 80 cm H₂o, respiratory rate I/E ratio. Technical Specification : Peak flow rate in CMV-ACMV:100/min. Peak flow rate in PS-PEEP:130L/min. Gas consumption : 1L/min. Pneumatic supply : 2.8 to 6bar Internal electrical power supply: NiCd batteries pack. Battery life : 10h, charging time : 10h, Main power supply : Input : 230v/50hz. Output :15v/6w. Dimensions (LxHxD) 250x210x170 mm Weight : 5.0kg. Accessories : patient circuit including disposable pipe, control pipe, three way valve, bracelets, adult mask, child's mask, carrying case, NiCd batteries, charger.</p>	
7.	<p>Defibrillator / Monitor :</p> <p><5S to charge to 360j with battery, 100 charger/discharge with fully charged new battery, event recording, thermal printing, LCD for sharp bright non fade display. Machine supply with following standard accessories. Defibrillator-paddles (adult & pediatric) 1 No., Disposable electrodes : 30 Nos, Patient cable (3 lead) 1 No., Mains cable (Power cord) 1 No., internal paddles: 1 No., Card jelly : 1 bottle, Earth cable 1 No., Accessory bag : 1 No., User manual : 1 No., Recording paper : 1 Roll – 20 mtrs.</p>	

8.	<p>Multipara monitor with 3 lead ECG, SPO2, Pulse, NICP:-</p> <p>High resolution 10.4" TFT color display. Standard configuration of ECG, Resp. Spo2, NIBP, Temp. Drug Dose calculation/HRV Analysis / Oxy CRG. Arrhythmia detection. 90 days trend . 3 lead / 5 lead monitoring. Pacemaker detection function. Built in three channel thermal recorder – optional. Display of 8 waveforms, including 4 ECG waveforms display, Customised waveforms display order</p>	
9.	<p>Oxygen therapy mask:-</p> <p>Ambo bag with face mask silicone</p>	
10.	<p>Laryngoscope set along with cells:-</p> <p>Macintosh with four blades - adult Fiber optic – pediatric.</p>	
11.	<p>Neck immobilizer (Hard Cervical Collar, Small & Medium):-</p> <p>Should be of exclusive design and shape, tested by EMS on the road allows to complete the head immobilization. The contour is to be realized with moulded antitoxic PVC and polyurethane, which will not absorb blood or any other body fluids. If added to a spinal board with its Velcro straps and to a cervical collar allows to move a patient even in very ruined areas, then you can keep it on the patient in X-ray chamber due to its adiolucent materials. Ideal for adult and child. Complete supply to include immobilizer base, removable head pad, two foam blocks and head straps.</p>	
12.	<p>Splints inflatable for immobilizer of limbs:-</p> <p>Vacuum Splint kit includes one arm (can be used to immobilize leg injures of children as well as lower leg and foot injures of adults), one leg splint (either the arm or leg splint may be used as full body immobilizer for infants and small children), one ankle splint, one foot vacuum pump all inside a orange PVC carrying case.</p>	
13.	<p>Oxygen cylinder</p>	
14.	<p>Emergency kit :-</p> <p>Emergency Kit Safety bag with 2 small side pockets and hard plastic case (removable) with Velcro lock & shoulder strap Sphygmomanometer - 1 Stethoscope - 1 Oxygen bottle 0.5L - 1 Oxygen pressure reducer with diaphragm fixed calibration - 1 Connecting tube - 1 Resuscitation Bag - 1 Adult mask V - 1 Paediatric mask III - 1 Portable manual aspirator - 1 Guedel airway set (1,2,3) - 1 Magill Forceps - 1 Universal scissor - 1 Tongue forceps - 1 Tourniquet – 1 Plastic penlight - 1 Emergency blanket – 1</p>	

15.	<p>Spine Board :-</p> <p>The spine board is to be built from high density cross linked polyethylene - this avoids splintering and cracking. The surface is to be impervious to body fluids and petroleum products being easy to keep clean. Its features to include seamless moulding to eliminate ingress of fluid, firm surface for CPR & immobilization. Compact dimensions for easy manoeuvring, 40.5X181 cms ,thickness 4.5 cms. Provision for positions for cervical collars or head immobilizers. Easy underside to allow easy lifting access, compatibility with standard rescue basket stretcher. The spine board is x-ray translucent and is extremely important for spinal injury casualties. Weight : less than 5 kgs. And load capacity more than 170 kg.</p>	
16.	<p>Fracture splint kit :-</p> <p>Flexible splints made from blue washable neoprene with ambular inlay and wide velcro tapes for fixation. Set of 5 pieces include:Arm, Wrist,Full leg,Full Arm Straight, Angular arm supplied in orange nylon bag with zip. Set's weight only 1.2 kgs. Set's size :67cm x 30 cm x 5 cm.</p>	
17.	<p>Syringe infusion pump (Infutek 400) :-</p> <p>PCA upgradable syringe pump accepting 10,20,30,50 ml syringes of known brands as well as custom programmable syringe with facility for time locked BOLUS/PCA facility and administration counter and display for the same. Body weight mode for automatic flow rate computation depending on patient's weight. Minimum six stage programming of target time and target volume. Programmable rate KVO (keep vein open), LED display of flow rate & LCD display of drug name, infused volume, infusion time simultaneously, facility to change flow rates in run condition. Facility to display name of the medicine being infused out of a library of 48 medicine.</p>	

Seal

Signature and Seal of Tenderer

Date